

## Human MuV-IgM (Mumps Virus-Immunoglobulin M) ELISA Kit

**Catalogue No.:** EH4440

**Size:** 96T

**Reactivity:** Human

**Application:** This immunoassay kit allows for the qualitative determination of MuV-IgM in human serum.

**Storage:** 2-8°C

**Expiry Date:** see kit label

**Principle:** Capture

**NOTE: FOR RESEARCH USE ONLY.**

### Kit Components

Item	Specifications(96T)	Storage
Micro ELISA Plate(Dismountable)	12 × 8	2-8°C /-20°C
MuV-IgM Positive Control	0.5ml×1	2-8°C
MuV-IgM Negative Control	0.5ml×1	2-8°C
Sample Dilution Buffer	12ml×1	2-8°C
HRP- Conjugates	12ml×1	2-8°C
TMB substrate A	6ml×1	2-8°C (protect from light)
TMB substrate B	6ml×1	2-8°C (protect from light)
Stop solution	6ml×1	2-8°C
Wash buffer (20X)	50ml×1	2-8°C
Plate Sealer	3pieces	
Product Description	1 copy	

## Principle of the Assay

This kit was based on Capture ELISA. Anti-Human-IgM ( $\mu$ chain) was pre-coated onto 96-well plates. The test samples were added to the wells, unbound conjugates were washed away with wash buffer. Then added HRP-Conjugates, if there were any MuV-IgM in the samples, it would form a Anti-Human-IgM ( $\mu$ chain) - MuV-IgM-HRP- MuV-Ag complex. TMB substrates were used to visualize HRP enzymatic reaction. It was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The optical density of developed color is read with a suitable photometer at 450nm with a selected reference wavelength within 650 nm.

## Precautions for Use

1. To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
2. After opening and before using, keep plate dry.
3. Before using the kit, spin tubes and bring down all components to the bottom of tubes.
4. Storage TMB reagents avoid light.
5. Washing process is very important, not fully wash easily cause a false positive and high background.
6. Duplicate well assay is recommended for both standard and sample testing.
7. Don't let microplate dry at the assay, for dry plate will inactivate active components on plate.
8. Don't reuse tips and tubes to avoid cross contamination.
9. Please do not mix the reagents in different kits of our company. Do not mix reagents from other manufacturers.
10. To ensure accurate results, proper adhesion of plate sealers during incubation steps is necessary.

## Material Required But Not Supplied

1. Microplate reader (wavelength: 450nm)
2. 37°C incubator
3. Automated plate washer
4. Precision single and multi-channel pipette and disposable tips
5. Clean tubes and Eppendorf tubes
6. Deionized or distilled water

## Manual Washing

Discard the solution in the plate without touching the sidewalls. Clap the plate on absorbent filter papers or other absorbent material. Fill each well completely with 350ul wash buffer and soak for 1 to 2 minutes, then

aspirate contents from the plate, and clap the plate on absorbent filter papers or other absorbent material. Repeat this procedure two more times for a total of THREE washes.

## Automated Washing

Aspirate all wells, wash plate THREE times with 350ul wash buffer. After the final wash, invert plate, and clap the plate on absorbent filter papers or other absorbent material. It is recommended that the washer be set for a soaking time of 1 minute.

## Sample Collection and Storage

Isolate the test samples soon after collecting, then, analyze immediately (within 2 hours). Or aliquot and store at  $-20^{\circ}\text{C}$  for long term. Avoid multiple freeze-thaw cycles.

- Serum: Coagulate the serum at room temperature (about 1 hour). Centrifuge at approximately  $1000 \times g$  for 15 min. Analyze the serum immediately or aliquot and store at  $-20^{\circ}\text{C}$ .

Note: Samples to be used within 3 days may be stored at  $2-8^{\circ}\text{C}$ , otherwise samples must be stored at  $-20^{\circ}\text{C}$  ( $\leq 1$  month) or  $-80^{\circ}\text{C}$  ( $\leq 2$  months) to avoid loss of bioactivity and contamination. Hemolyzed samples are not suitable for use in this assay.

## Wash Buffer Preparation:

Dilute 50mL of Concentrated Wash Buffer to 1000 mL of Wash Buffer with deionized or distilled water.

## Assay Procedure

1. Bring all reagents to room temperature for 30 min before use.
2. Label the sample wells, 3 Negative Controls, 2 Positive Control and 1 blank well.
3. Add  $100\mu\text{L}$  Negative Controls and Positive Controls to each well (except blank well).
4. Add 100ul sample dilution buffer to each sample well (except blank well). Then Add 10ul sample to each well and gently tap the plate to ensure thorough mixing. Seal the plate with a cover and incubate at  $37^{\circ}\text{C}$  for 30 min.
5. Remove the cover, and wash plate 5 times with Wash buffer and let the wash buffer stay in the wells for 1 minute each time.

6. Add 100 µL HRP- Conjugates into each well (except blank well) and gently tap the plate to ensure thorough mixing.
7. Seal the plate with a cover and incubate at 37°C for 30 min.
8. Remove the cover, and wash plate 5 times with Wash buffer and let the wash buffer stay in the wells for 1 minute each time.
9. Add 50 µl of TMB substrate A and 50 µl of TMB substrate B into each well. Gently tap the plate to ensure thorough mixing. Cover the plate and incubate at 37°C in dark for 15 min. And the shades of blue can be seen in the Positive Controls. Negative Controlswells show no obvious color.
10. Add 50 µl of Stop solution into each well and mix thoroughly. The color changes into yellow immediately.
11. Read the O.D. absorbance at 450 nm in a microplate reader immediately after adding the stop solution. (Use the blank well to set zero).

### 1. Calculation of the Cutoff Value

Cutoff Value =  $NCx+0.1$

NCx: Mean Absorbance of Negative Control(When the negative mean A value is less than 0.05, it is calculated as 0.05.When the negative mean A value is greater than or equal to 0.05, it is calculated according to the actual value.)

### 2. Determination of results

Sample with absorbance values  $\leq$  Cutoff Value are NON-REACTIVE and are considered NEGATIVE for MuV-IgM. Sample with absorbance values  $>$  Cutoff Value are considered POSITIVE for MuV-IgM.

### 3. Quality control

The blank well (only adding TMB and Stop solution) should not be greater than 0.08.

The positive control (PC) A value was greater than 0.80.

The negative mean A value was less than 0.10.